

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120925\
 Data File : VX048769.D
 Acq On : 09 Dec 2025 09:46
 Operator : JC/MD
 Sample : VX1209WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1209WBL01

Quant Time: Dec 10 00:13:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	129302	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	266304	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	288359	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	149395	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	97271	56.049	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.100%
35) Dibromofluoromethane	5.373	113	83404	45.486	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.980%
50) Toluene-d8	8.635	98	299705	46.630	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.260%
62) 4-Bromofluorobenzene	11.061	95	127733	57.632	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	115.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120925\
Data File : VX048769.D
Acq On : 09 Dec 2025 09:46
Operator : JC/MD
Sample : VX1209WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1209WBL01

Quant Time: Dec 10 00:13:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

