



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX120925\  
Data File : VX048767.D  
Acq On : 09 Dec 2025 08:16  
Operator : JC/MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X120425W.M  
Title : SW846 8260  
Last Update : Fri Dec 05 03:27:34 2025



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1629

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 19.2         | 3982       | PASS                |
| 75             | 95              | 30              | 60              | 48.0         | 9980       | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 20771      | PASS                |
| 96             | 95              | 5               | 9               | 6.7          | 1393       | PASS                |
| 173            | 174             | 0.00            | 2               | 1.3          | 212        | PASS                |
| 174            | 95              | 50              | 100             | 76.9         | 15981      | PASS                |
| 175            | 174             | 5               | 9               | 7.6          | 1216       | PASS                |
| 176            | 174             | 95              | 101             | 97.2         | 15532      | PASS                |
| 177            | 176             | 5               | 9               | 6.2          | 968        | PASS                |