

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121018\
 Data File : VX006375.D
 Acq On : 10 Dec 2018 17:32
 Operator : JC/MD
 Sample : J6288-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW30212-20181205

Quant Time: Dec 11 04:35:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	768385	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1174241	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1057367	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	510309	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	380828	48.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.10%	
35) Dibromofluoromethane	5.50	113	366131	48.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.32%	
50) Toluene-d8	8.71	98	1360014	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
62) 4-Bromofluorobenzene	11.13	95	501223	48.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.20%	
Target Compounds						
44) Trichloroethene	7.21	130	24722	2.755	ug/l	93
64) Tetrachloroethene	9.33	164	2892	0.363	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121018\
Data File : VX006375.D
Acq On : 10 Dec 2018 17:32
Operator : JC/MD
Sample : J6288-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW30212-20181205

Quant Time: Dec 11 04:35:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

