

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121018\
 Data File : VX006377.D
 Acq On : 10 Dec 2018 18:25
 Operator : JC/MD
 Sample : J6288-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW301S-20181206

Quant Time: Dec 11 04:39:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	744731	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1142337	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1037279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	495301	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	380841	50.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.18%	
35) Dibromofluoromethane	5.51	113	356869	48.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.50%	
50) Toluene-d8	8.71	98	1329437	49.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.84%	
62) 4-Bromofluorobenzene	11.13	95	493815	48.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.42%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121018\
Data File : VX006377.D
Acq On : 10 Dec 2018 18:25
Operator : JC/MD
Sample : J6288-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW301S-20181206

Quant Time: Dec 11 04:39:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

