

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121018\
 Data File : VX006392.D
 Acq On : 11 Dec 2018 01:58
 Operator : JC/MD
 Sample : J6324-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-103D3-20181205

Quant Time: Dec 11 05:55:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	698582	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1105974	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	993346	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	479182	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	365952	51.32	ug/l	0.00
Spiked Amount			Recovery	=	102.64%	
35) Dibromofluoromethane	5.51	113	337136	47.08	ug/l	0.00
Spiked Amount			Recovery	=	94.16%	
50) Toluene-d8	8.71	98	1276504	49.01	ug/l	0.00
Spiked Amount			Recovery	=	98.02%	
62) 4-Bromofluorobenzene	11.13	95	477278	48.63	ug/l	0.00
Spiked Amount			Recovery	=	97.26%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.39	101	11618	1.759	ug/l	92
27) cis-1,2-Dichloroethene	4.61	96	4287	0.571	ug/l	83
30) Chloroform	5.21	83	4547	0.388	ug/l	92
44) Trichloroethene	7.21	130	2674361	316.475	ug/l	93
64) Tetrachloroethene	9.34	164	4285	0.573	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121018\
Data File : VX006392.D
Acq On : 11 Dec 2018 01:58
Operator : JC/MD
Sample : J6324-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-103D3-20181205

Quant Time: Dec 11 05:55:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

