

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121018\
 Data File : VX006393.D
 Acq On : 11 Dec 2018 02:25
 Operator : JC/MD
 Sample : J6324-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-103D2-20181205

Quant Time: Dec 11 05:57:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	713849	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1128779	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1018279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	486990	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	374005	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
35) Dibromofluoromethane	5.51	113	347218	47.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.02%	
50) Toluene-d8	8.71	98	1315530	49.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.98%	
62) 4-Bromofluorobenzene	11.13	95	489235	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	12042	1.785	ug/l	95
27) cis-1,2-Dichloroethene	4.60	96	4693	0.611	ug/l	87
30) Chloroform	5.22	83	7268	0.607	ug/l	89
44) Trichloroethene	7.21	130	3682132	426.928	ug/l	94
64) Tetrachloroethene	9.34	164	5425	0.707	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121018\
Data File : VX006393.D
Acq On : 11 Dec 2018 02:25
Operator : JC/MD
Sample : J6324-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-103D2-20181205

Quant Time: Dec 11 05:57:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

