

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121018\
 Data File : VX006398.D
 Acq On : 11 Dec 2018 04:38
 Operator : JC/MD
 Sample : J6324-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-109D3-20181206

Quant Time: Dec 11 06:08:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	682595	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1076404	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	989353	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	469261	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	360509	51.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.48%	
35) Dibromofluoromethane	5.51	113	329049	47.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.42%	
50) Toluene-d8	8.71	98	1266361	49.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.90%	
62) 4-Bromofluorobenzene	11.14	95	470598	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.39	101	16079	2.492	ug/l	97
27) cis-1,2-Dichloroethene	4.60	96	6513	0.887	ug/l	90
38) Carbon Tetrachloride	5.78	117	6085	0.571	ug/l	86
44) Trichloroethene	7.21	130	487843	59.316	ug/l	96
64) Tetrachloroethene	9.34	164	5349	0.718	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121018\
Data File : VX006398.D
Acq On : 11 Dec 2018 04:38
Operator : JC/MD
Sample : J6324-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-109D3-20181206

Quant Time: Dec 11 06:08:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

