

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121018\
 Data File : VX006399.D
 Acq On : 11 Dec 2018 05:05
 Operator : JC/MD
 Sample : J6324-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-109D2-20181206

Quant Time: Dec 11 09:38:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	686600	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1079797	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	984886	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	468237	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	364660	52.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.06%	
35) Dibromofluoromethane	5.51	113	332642	47.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.16%	
50) Toluene-d8	8.71	98	1264878	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
62) 4-Bromofluorobenzene	11.13	95	477644	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	7620	1.174	ug/l	96
44) Trichloroethene	7.21	130	278868	33.800	ug/l	95
49) 1,4-Dioxane	7.75	88	2312	11.357	ug/l	90
64) Tetrachloroethene	9.34	164	3239	0.437	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121018\
Data File : VX006399.D
Acq On : 11 Dec 2018 05:05
Operator : JC/MD
Sample : J6324-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-109D2-20181206

Quant Time: Dec 11 09:38:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

