

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121018\
 Data File : VX006400.D
 Acq On : 11 Dec 2018 05:32
 Operator : JC/MD
 Sample : J6324-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-104D1-20181206

Quant Time: Dec 11 09:39:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	683555	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1084021	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	983598	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	472656	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	365466	52.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.74%	
35) Dibromofluoromethane	5.50	113	333913	47.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.16%	
50) Toluene-d8	8.71	98	1265684	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
62) 4-Bromofluorobenzene	11.14	95	473296	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	8380	1.297	ug/l #	94
44) Trichloroethene	7.21	130	383123	46.256	ug/l	97
49) 1,4-Dioxane	7.75	88	2570	12.575	ug/l #	89
64) Tetrachloroethene	9.34	164	16152	2.181	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121018\
Data File : VX006400.D
Acq On : 11 Dec 2018 05:32
Operator : JC/MD
Sample : J6324-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-104D1-20181206

Quant Time: Dec 11 09:39:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

