

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121018\
 Data File : VX006406.D
 Acq On : 11 Dec 2018 08:12
 Operator : JC/MD
 Sample : J6314-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampleId :
 BPS1-TT-MW3011-20181207

Quant Time: Dec 11 09:47:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	664098	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1059368	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	961452	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	452032	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	361956	53.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.78%	
35) Dibromofluoromethane	5.51	113	326709	47.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.26%	
50) Toluene-d8	8.71	98	1239282	49.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.34%	
62) 4-Bromofluorobenzene	11.13	95	463622	49.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.62%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121018\
 Data File : VX006406.D
 Acq On : 11 Dec 2018 08:12
 Operator : JC/MD
 Sample : J6314-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampled :
 BPS1-TT-MW301I-20181207

Quant Time: Dec 11 09:47:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

