

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121018\
 Data File : VX006407.D
 Acq On : 11 Dec 2018 08:38
 Operator : JC/MD
 Sample : J6314-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW30311-20181207

Quant Time: Dec 11 09:47:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	659762	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1048052	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	947873	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	437031	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	352706	52.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.74%	
35) Dibromofluoromethane	5.51	113	324334	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
50) Toluene-d8	8.71	98	1227351	49.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.46%	
62) 4-Bromofluorobenzene	11.13	95	448874	48.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.52%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	6001	0.846	ug/l	94
44) Trichloroethene	7.21	130	42046	5.251	ug/l	94
64) Tetrachloroethene	9.33	164	139082	19.484	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121018\
Data File : VX006407.D
Acq On : 11 Dec 2018 08:38
Operator : JC/MD
Sample : J6314-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW30311-20181207

Quant Time: Dec 11 09:47:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

