

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121018\
 Data File : VX006368.D
 Acq On : 10 Dec 2018 14:25
 Operator : JC/MD
 Sample : J6332-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Dec 11 04:24:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	843225	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1249114	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1133960	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	547768	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	410252	47.66	ug/l	0.00
Spiked Amount			Recovery	=	95.32%	
35) Dibromofluoromethane	5.50	113	389923	48.21	ug/l	0.00
Spiked Amount			Recovery	=	96.42%	
50) Toluene-d8	8.71	98	1453097	49.40	ug/l	0.00
Spiked Amount			Recovery	=	98.80%	
62) 4-Bromofluorobenzene	11.13	95	537581	48.49	ug/l	0.00
Spiked Amount			Recovery	=	96.98%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121018\
Data File : VX006368.D
Acq On : 10 Dec 2018 14:25
Operator : JC/MD
Sample : J6332-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Dec 11 04:24:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.00 min

Lab File: VX006368.D

Acq: 10 Dec 2018 14:25

Instrument :

MSVOA_X

ClientSampleId :


STORAGE-BLANK-SOIL-REF-6

Tot Ion:168 Resp: 843225

Ion Ratio Lower Upper

168 100

99 49.7 37.8 56.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67



300000

200000

100000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 47.660 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.00 min

Lab File: VX006368.D


Acq: 10 Dec 2018 14:25

Tgt Ion: 65 Resp: 410252

Ion Ratio Lower Upper

65 100

67 52.6 0.0 107.2



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07


150000


100000

50000

0

Time--> 5.90 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. 0.00 min

Lab File: VX006368.D

Acq: 10 Dec 2018 14:25

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-SOIL-REF-6


Tot Ion:114 Resp: 1249114

Ion Ratio Lower Upper

114 100

63 17.4 0.0 36.6




88 14.9 0.0 29.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 48.213 ug/l

RT: 5.50 min Scan# 713

Delta R.T. 0.00 min

Lab File: VX006368.D

Acq: 10 Dec 2018 14:25


Tgt Ion:113 Resp: 389923

Ion Ratio Lower Upper

113 100

111 102.3 81.1 121.7



192 22.0 18.2 27.2



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.395 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX006368.D

Acq: 10 Dec 2018 14:25

Instrument :

MSVOA_X

ClientSampleId :


STORAGE-BLANK-SOIL-REF-6

Tot Ion: 98 Resp: 1453097

Ion Ratio Lower Upper




98 100

100 64.7 52.2 78.2



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V



#62

4-Bromofluorobenzene

Concen: 48.494 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.01 min

Lab File: VX006368.D

Acq: 10 Dec 2018 14:25


Tgt Ion: 95 Resp: 537581

Ion Ratio Lower Upper

95 100

174 89.5 0.0 187.0



176 86.9 0.0 181.8



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. 0.00 min
Lab File: VX006368.D
Acq: 10 Dec 2018 14:25

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Tot Ion:117 Resp: 1133960
Ion Ratio Lower Upper
117 100
82 49.0 39.6 59.4
119 32.0 26.3 39.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX006368.D
Acq: 10 Dec 2018 14:25

Tgt Ion:152 Resp: 547768
Ion Ratio Lower Upper
152 100
115 56.3 39.0 116.9
150 157.5 0.0 345.8

