

Data File : VX013920.D

Acq On : 11 Dec 2019 00:33

Operator : JC/SP

Sample : K6185-07DL 5X

Misc : 5.0mL/MSVOA_X/WATER

ALS Vial : 37 Sample Multiplier: 1

Instrument :

MSVOA_X

ClientSampleId :

RE131D2-20191205DL

Quant Time: Dec 11 04:15:07 2019

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M

Quant Title : SW846 8260

QLast Update : Tue Nov 26 15:53:00 2019

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	590092	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	906203	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	802104	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	351362	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.05	65	334525	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
35) Dibromofluoromethane	5.49	113	271826	48.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.04%	
50) Toluene-d8	8.71	98	1071353	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
62) 4-Bromofluorobenzene	11.13	95	356324	45.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.54%	

Target Compounds

						Qvalue
5) Bromomethane	1.64	94	2433	0.437	ug/l #	60
9) 1,1,2-Trichlorotrifluoroet	2.37	101	94446	16.787	ug/l	98
11) Tert butyl alcohol	3.00	59	5641	3.336	ug/l #	83
13) Acrolein	2.28	56	218	6.192	ug/l #	15
16) Acetone	2.44	43	8689	2.569	ug/l	92
20) Methylene Chloride	2.84	84	2797	0.431	ug/l #	88
21) trans-1,2-Dichloroethene	3.16	96	1564	0.256	ug/l #	30
25) 2-Butanone	4.67	43	2144	0.427	ug/l #	80
27) cis-1,2-Dichloroethene	4.59	96	7636	1.087	ug/l	82
36) 1,1-Dichloropropene	5.65	75	41955	5.100	ug/l #	51
37) Ethyl Acetate	4.67	43	2144	0.232	ug/l	97
38) Carbon Tetrachloride	5.65	117	51796	6.829	ug/l #	15
44) Trichloroethene	7.20	130	129167	18.785	ug/l	100
49) 1,4-Dioxane	7.75	88	857	5.187	ug/l #	1
64) Tetrachloroethene	9.33	164	14963	2.502	ug/l	89
76) 1,2,3-Trichloropropane	11.14	75	173374	21.938	ug/l #	39
81) trans-1,4-Dichloro-2-buten	11.14	75	173374	60.020	ug/l #	11
92) 1,2-Dibromo-3-Chloropropan	12.89	75	413	0.222	ug/l	73

(#)=qualifier out of range (m)=manual integration (+)=signals summed

Data File : VX013920.D

Acq On : 11 Dec 2019 00:33

Operator : JC/SP

Sample : K6185-07DL 5X

Misc : 5.0mL/MSVOA_X/WATER

ALS Vial : 37 Sample Multiplier: 1

Instrument :

MSVOA_X

ClientSampleId :

RE131D2-20191205DL

Quant Time: Dec 11 04:15:07 2019

Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M

Quant Title : SW846 8260

QLast Update : Tue Nov 26 15:53:00 2019

Response via : Initial Calibration

