

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121019\
 Data File : VX013896.D
 Acq On : 10 Dec 2019 14:26
 Operator : JC/SP
 Sample : K6186-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-GAC2-IDW01

Quant Time: Dec 11 06:21:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	587469	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	892499	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	805104	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	356143	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	328493	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
35) Dibromofluoromethane	5.49	113	272159	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
50) Toluene-d8	8.71	98	1068046	49.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.36%	
62) 4-Bromofluorobenzene	11.13	95	358982	46.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.62%	
Target Compounds						
16) Acetone	2.43	43	19749	5.866	ug/l	97
19) Methyl tert-butyl Ether	3.19	73	8865	0.479	ug/l	96
27) cis-1,2-Dichloroethene	4.59	96	67994	9.721	ug/l	90
43) Isopropyl Acetate	6.43	43	19552	1.307	ug/l	96
44) Trichloroethene	7.21	130	3754	0.554	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121019\
Data File : VX013896.D
Acq On : 10 Dec 2019 14:26
Operator : JC/SP
Sample : K6186-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-GAC2-IDW01

Quant Time: Dec 11 06:21:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

