

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121019\
 Data File : VX013902.D
 Acq On : 10 Dec 2019 16:46
 Operator : JC/SP
 Sample : K6185-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D3-20191204

Quant Time: Dec 11 03:22:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	576511	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	887326	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	798119	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	342027	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	323472	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
35) Dibromofluoromethane	5.49	113	267830	48.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.64%	
50) Toluene-d8	8.71	98	1053265	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
62) 4-Bromofluorobenzene	11.14	95	350377	45.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.92%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluoroet	2.37	101	3412691	620.879	ug/l	99
12) 1,1-Dichloroethene	2.37	96	20609	3.756	ug/l	91
16) Acetone	2.43	43	9976	3.019	ug/l #	85
27) cis-1,2-Dichloroethene	4.59	96	28550	4.159	ug/l	87
44) Trichloroethene	7.21	130	1333078	198.000	ug/l	99
49) 1,4-Dioxane	7.74	88	5217	32.245	ug/l #	87
64) Tetrachloroethene	9.33	164	561839	94.397	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121019\
Data File : VX013902.D
Acq On : 10 Dec 2019 16:46
Operator : JC/SP
Sample : K6185-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20191204

Quant Time: Dec 11 03:22:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

