

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121019\
 Data File : VX013908.D
 Acq On : 10 Dec 2019 19:07
 Operator : JC/SP
 Sample : K6196-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-1

Manual Integrations
 APPROVED

apatel
 12/11/2019 2:14:09 PM

Quant Time: Dec 11 06:35:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	577031	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	893830	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	795647	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	347615	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	325692	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
35) Dibromofluoromethane	5.49	113	269372	48.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.50%	
50) Toluene-d8	8.71	98	1062479	49.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.70%	
62) 4-Bromofluorobenzene	11.14	95	350451	45.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.28%	
Target Compounds						
16) Acetone	2.43	43	5949m	1.799	ug/l	
27) cis-1,2-Dichloroethene	4.59	96	3817	0.556	ug/l	86
30) Chloroform	5.20	83	84675	7.959	ug/l	96
44) Trichloroethene	7.21	130	6001	0.885	ug/l	80
47) Bromodichloromethane	7.90	83	5220	0.658	ug/l #	90
64) Tetrachloroethene	9.33	164	169822	28.621	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121019\
Data File : VX013908.D
Acq On : 10 Dec 2019 19:07
Operator : JC/SP
Sample : K6196-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-1

Manual Integrations
APPROVED

apatel
12/11/2019 2:14:09 PM

Quant Time: Dec 11 06:35:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

