

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121019\
 Data File : VX013916.D
 Acq On : 10 Dec 2019 23:00
 Operator : JC/SP
 Sample : VX1210WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1210WBL02

Quant Time: Dec 11 04:36:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	578490	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	886414	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	781207	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	349416	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	321526	47.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.76%	
35) Dibromofluoromethane	5.49	113	264841	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
50) Toluene-d8	8.71	98	1040885	48.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.50%	
62) 4-Bromofluorobenzene	11.14	95	344960	44.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.62%	
Target Compounds						
16) Acetone	2.43	43	6684	2.016	ug/l #	88
44) Trichloroethene	7.20	130	3200	0.476	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121019\
Data File : VX013916.D
Acq On : 10 Dec 2019 23:00
Operator : JC/SP
Sample : VX1210WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1210WBL02

Quant Time: Dec 11 04:36:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

