

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121019\
 Data File : VX013919.D
 Acq On : 11 Dec 2019 00:10
 Operator : JC/SP
 Sample : K6185-06DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D1-20191205DL

Quant Time: Dec 11 06:48:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	552926	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	833970	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	749589	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	309117	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	309194	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
35) Dibromofluoromethane	5.49	113	254525	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
50) Toluene-d8	8.71	98	990462	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
62) 4-Bromofluorobenzene	11.14	95	328057	45.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.58%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.38	101	4234	0.803	ug/l	90
27) cis-1,2-Dichloroethene	4.59	96	7382	1.121	ug/l	86
44) Trichloroethene	7.21	130	236965	37.448	ug/l	92
64) Tetrachloroethene	9.33	164	15203	2.720	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121019\
Data File : VX013919.D
Acq On : 11 Dec 2019 00:10
Operator : JC/SP
Sample : K6185-06DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D1-20191205DL

Quant Time: Dec 11 06:48:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

