

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121019\
 Data File : VX013920.D
 Acq On : 11 Dec 2019 00:33
 Operator : JC/SP
 Sample : K6185-07DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D2-20191205DL

Quant Time: Dec 11 06:49:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	590092	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	906203	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	802104	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	351362	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	334525	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
35) Dibromofluoromethane	5.49	113	271826	48.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.04%	
50) Toluene-d8	8.71	98	1071353	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
62) 4-Bromofluorobenzene	11.13	95	356324	45.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.54%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	94446	16.787	ug/l	98
27) cis-1,2-Dichloroethene	4.59	96	7636	1.087	ug/l	82
44) Trichloroethene	7.20	130	129167	18.785	ug/l	100
64) Tetrachloroethene	9.33	164	14963	2.502	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121019\
Data File : VX013920.D
Acq On : 11 Dec 2019 00:33
Operator : JC/SP
Sample : K6185-07DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D2-20191205DL

Quant Time: Dec 11 06:49:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

