

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121019\
 Data File : VX013926.D
 Acq On : 11 Dec 2019 03:27
 Operator : JC/SP
 Sample : K6237-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS037-191209

Quant Time: Dec 11 07:18:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	578800	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	886109	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	807702	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	375536	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	324244	48.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.52%	
35) Dibromofluoromethane	5.49	113	269159	49.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.26%	
50) Toluene-d8	8.71	98	1047859	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
62) 4-Bromofluorobenzene	11.13	95	362267	47.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.14%	
Target Compounds						
16) Acetone	2.43	43	30953	9.332	ug/l	100
27) cis-1,2-Dichloroethene	4.59	96	32946	4.781	ug/l	88
44) Trichloroethene	7.22	130	7031	1.046	ug/l	81
65) Chlorobenzene	10.14	112	7198	0.429	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121019\
Data File : VX013926.D
Acq On : 11 Dec 2019 03:27
Operator : JC/SP
Sample : K6237-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS037-191209

Quant Time: Dec 11 07:18:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

