


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX121019\
Data File : VX013886.D
Acq On : 10 Dec 2019 09:52
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Title : SW846 8260
Last Update : Tue Nov 26 15:53:00 2019



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1673

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	25067	PASS
75	95	30	60	48.1	64397	PASS
95	95	100	100	100.0	133888	PASS
96	95	5	9	6.6	8825	PASS
173	174	0.00	2	1.1	1240	PASS
174	95	50	100	81.6	109196	PASS
175	174	5	9	7.9	8621	PASS
176	174	95	101	100.0	109189	PASS
177	176	5	9	6.3	6927	PASS