

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019812.D
 Acq On : 10 Dec 2020 09:54
 Operator : JC/MD
 Sample : VX1210WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1210WBL01

Quant Time: Dec 11 01:03:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	324571	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.82	114	535941	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	480847	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	204439	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	201249	47.18	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.36%
35) Dibromofluoromethane	5.46	113	162963	47.05	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.10%
50) Toluene-d8	8.69	98	643234	48.64	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.28%
62) 4-Bromofluorobenzene	11.12	95	222331	44.17	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.34%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019812.D
Acq On : 10 Dec 2020 09:54
Operator : JC/MD
Sample : VX1210WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1210WBL01

Quant Time: Dec 11 01:03:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

