

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019819.D
 Acq On : 10 Dec 2020 12:47
 Operator : JC/MD
 Sample : L4996-25
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB01-20201208

Quant Time: Dec 11 01:25:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	319822	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	533555	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	483494	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	209742	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	202355	48.14	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.28%
35) Dibromofluoromethane	5.46	113	162627	47.16	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.32%
50) Toluene-d8	8.70	98	641647	48.74	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.48%
62) 4-Bromofluorobenzene	11.12	95	222296	44.36	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.72%

Target Compounds					Qvalue
16) Acetone	2.42	43	8161	5.165 ug/l	100
25) 2-Butanone	4.65	43	1610	0.663 ug/l	96
29) Tetrahydrofuran	5.10	42	1223	0.782 ug/l #	82
52) Toluene	8.76	92	24007	2.606 ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019819.D
Acq On : 10 Dec 2020 12:47
Operator : JC/MD
Sample : L4996-25
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB01-20201208

Quant Time: Dec 11 01:25:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

