

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019820.D
 Acq On : 10 Dec 2020 13:11
 Operator : JC/MD
 Sample : L4996-26
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01-20201208

Quant Time: Dec 11 01:27:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	309322	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	524117	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	476691	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	204257	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	199597	49.10	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.20%
35) Dibromofluoromethane	5.46	113	160295	47.32	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.64%
50) Toluene-d8	8.70	98	633112	48.96	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.92%
62) 4-Bromofluorobenzene	11.12	95	218466	44.39	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.78%

Target Compounds

					Qvalue
3) Chloromethane	1.31	50	954	0.286	ug/l 99
16) Acetone	2.42	43	8698	5.692	ug/l 98
25) 2-Butanone	4.64	43	1734	0.738	ug/l # 89
29) Tetrahydrofuran	5.10	42	1250	0.826	ug/l # 86
52) Toluene	8.76	92	24163	2.670	ug/l 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019820.D
Acq On : 10 Dec 2020 13:11
Operator : JC/MD
Sample : L4996-26
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-20201208

Quant Time: Dec 11 01:27:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

