

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019821.D
 Acq On : 10 Dec 2020 13:34
 Operator : JC/MD
 Sample : L5020-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 201209001-02-TRIP-BLANK

Quant Time: Dec 11 01:29:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	312459	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	515488	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	474708	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	209295	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	196776	47.92	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.84%		
35) Dibromofluoromethane	5.46	113	159744	47.95	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.90%		
50) Toluene-d8	8.70	98	621674	48.88	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	97.76%		
62) 4-Bromofluorobenzene	11.12	95	220419	45.53	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.06%		
Target Compounds						
16) Acetone	2.42	43	3261	2.113	ug/l	99
30) Chloroform	5.17	83	2516	0.415	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019821.D
Acq On : 10 Dec 2020 13:34
Operator : JC/MD
Sample : L5020-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
201209001-02-TRIP-BLANK

Quant Time: Dec 11 01:29:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

