

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019824.D
 Acq On : 10 Dec 2020 14:44
 Operator : JC/MD
 Sample : L4996-05 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D2-20201208

Quant Time: Dec 11 12:41:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	295983	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	491764	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	456316	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	223440	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.02	65	184651	47.47	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.94%		
35) Dibromofluoromethane	5.46	113	147535	46.42	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	92.84%		
50) Toluene-d8	8.70	98	591305	48.73	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	97.46%		
62) 4-Bromofluorobenzene	11.12	95	223120	48.31	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	96.62%		
Target Compounds						
27) cis-1,2-Dichloroethene	4.56	96	1121	0.303	ug/l	77
44) Trichloroethene	7.18	130	524976	147.400	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019824.D
Acq On : 10 Dec 2020 14:44
Operator : JC/MD
Sample : L4996-05 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D2-20201208

Quant Time: Dec 11 12:41:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

