

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019826.D
 Acq On : 10 Dec 2020 15:30
 Operator : JC/MD
 Sample : L4996-13DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D4-20201207DL

Quant Time: Dec 11 01:59:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	327282	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	536401	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	478316	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	197636	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	202151	47.00	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.00%
35) Dibromofluoromethane	5.46	113	164193	47.36	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.72%
50) Toluene-d8	8.70	98	645449	48.77	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.54%
62) 4-Bromofluorobenzene	11.12	95	218243	43.32	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.64%

Target Compounds

					Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.37	101	4101	1.383	ug/l 96
12) 1,1-Dichloroethene	2.36	96	791	0.264	ug/l 84
27) cis-1,2-Dichloroethene	4.57	96	1873	0.458	ug/l 98
44) Trichloroethene	7.19	130	170404	43.864	ug/l 99
49) 1,4-Dioxane	7.73	88	410	4.325	ug/l # 86
64) Tetrachloroethene	9.32	164	3548	1.106	ug/l 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019826.D
Acq On : 10 Dec 2020 15:30
Operator : JC/MD
Sample : L4996-13DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20201207DL

Quant Time: Dec 11 01:59:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

