

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019827.D
 Acq On : 10 Dec 2020 15:54
 Operator : JC/MD
 Sample : L4996-07 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20201208

Quant Time: Dec 11 02:01:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	323019	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	534817	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	494770	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	214630	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	204513	48.17	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.34%
35) Dibromofluoromethane	5.46	113	164743	47.66	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.32%
50) Toluene-d8	8.70	98	644976	48.88	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.76%
62) 4-Bromofluorobenzene	11.12	95	227660	45.33	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.66%

Target Compounds					Qvalue
44) Trichloroethene	7.18	130	17971	4.640	ug/l 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019827.D
Acq On : 10 Dec 2020 15:54
Operator : JC/MD
Sample : L4996-07 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20201208

Quant Time: Dec 11 02:01:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

