

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019829.D
 Acq On : 10 Dec 2020 16:40
 Operator : JC/MD
 Sample : L4996-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE105D1-20201208

Quant Time: Dec 11 12:43:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	317818	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	528575	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	481834	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	207129	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	204318	48.91	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.82%
35) Dibromofluoromethane	5.46	113	162285	47.51	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.02%
50) Toluene-d8	8.70	98	639669	49.05	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.10%
62) 4-Bromofluorobenzene	11.12	95	224755	45.28	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.56%

Target Compounds

					Qvalue
2) Dichlorodifluoromethane	1.18	85	1358	0.467	ug/l 95
9) 1,1,2-Trichlorotrifluoroet	2.37	101	15916	5.526	ug/l 99
12) 1,1-Dichloroethene	2.36	96	2683	0.921	ug/l 92
27) cis-1,2-Dichloroethene	4.56	96	4674	1.176	ug/l 91
30) Chloroform	5.18	83	1322	0.214	ug/l 96
44) Trichloroethene	7.19	130	307827	80.411	ug/l 99
49) 1,4-Dioxane	7.71	88	1252	13.403	ug/l # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019829.D
Acq On : 10 Dec 2020 16:40
Operator : JC/MD
Sample : L4996-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE105D1-20201208

Quant Time: Dec 11 12:43:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

