

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121020\
 Data File : VX019830.D
 Acq On : 10 Dec 2020 17:03
 Operator : JC/MD
 Sample : L4996-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D1-20201208

Quant Time: Dec 11 00:50:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	315435	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	522450	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	473284	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	200374	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	200950	48.47	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.94%
35) Dibromofluoromethane	5.464	113	160565	47.55	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.10%
50) Toluene-d8	8.695	98	628227	48.74	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.48%
62) 4-Bromofluorobenzene	11.116	95	216816	44.19	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.38%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.367	101	1691	0.59	ug/l	Qvalue 100
44) Trichloroethene	7.184	130	133786	35.36	ug/l	99
49) 1,4-Dioxane	7.726	88	1135	12.29	ug/l #	89
64) Tetrachloroethene	9.317	164	5995	1.89	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121020\
Data File : VX019830.D
Acq On : 10 Dec 2020 17:03
Operator : JC/MD
Sample : L4996-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D1-20201208

Quant Time: Dec 11 00:50:25 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

