

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019840.D
 Acq On : 10 Dec 2020 22:05
 Operator : JC/MD
 Sample : VX1210WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1210WBL02

Quant Time: Dec 11 02:53:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	326857	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	536442	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	486471	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	205332	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	203220	47.31	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.62%		
35) Dibromofluoromethane	5.46	113	166411	48.00	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.00%		
50) Toluene-d8	8.70	98	650896	49.18	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.36%		
62) 4-Bromofluorobenzene	11.12	95	225310	44.72	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	89.44%		

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019840.D
Acq On : 10 Dec 2020 22:05
Operator : JC/MD
Sample : VX1210WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1210WBL02

Quant Time: Dec 11 02:53:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

