

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019843.D
 Acq On : 10 Dec 2020 23:15
 Operator : JC/MD
 Sample : PB133508TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133508TB

Quant Time: Dec 11 02:58:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	339203	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	562307	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	518124	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	222537	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	211121	47.36	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.72%		
35) Dibromofluoromethane	5.46	113	172766	47.54	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.08%		
50) Toluene-d8	8.70	98	682942	49.22	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.44%		
62) 4-Bromofluorobenzene	11.12	95	240321	45.51	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.02%		

Target Compounds

					Qvalue
16) Acetone	2.42	43	30014	17.911	ug/l 95
20) Methylene Chloride	2.83	84	70412	17.481	ug/l 98
43) Isopropyl Acetate	6.41	43	22141	2.655	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019843.D
Acq On : 10 Dec 2020 23:15
Operator : JC/MD
Sample : PB133508TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133508TB

Quant Time: Dec 11 02:58:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

