

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019844.D
 Acq On : 10 Dec 2020 23:38
 Operator : JC/MD
 Sample : L4975-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 21222

Quant Time: Dec 11 03:00:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	336089	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	554426	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	510581	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	218636	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	211554	47.89	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.78%		
35) Dibromofluoromethane	5.46	113	169610	47.34	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.68%		
50) Toluene-d8	8.70	98	678120	49.57	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	99.14%		
62) 4-Bromofluorobenzene	11.12	95	237618	45.64	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.28%		

Target Compounds					Qvalue
16) Acetone	2.42	43	39641	23.876	ug/l 97
18) Methyl Acetate	2.75	43	5152	1.268	ug/l 94
20) Methylene Chloride	2.83	84	31494	7.088	ug/l 99
43) Isopropyl Acetate	6.41	43	24444	2.972	ug/l 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019844.D
Acq On : 10 Dec 2020 23:38
Operator : JC/MD
Sample : L4975-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
21222

Quant Time: Dec 11 03:00:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

