

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019846.D
 Acq On : 11 Dec 2020 00:24
 Operator : JC/MD
 Sample : L4983-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR1-A-BOTTOM

Quant Time: Dec 11 03:12:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	323809	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	542014	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	506993	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	220521	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	206273	48.47	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.94%		
35) Dibromofluoromethane	5.46	113	166105	47.42	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.84%		
50) Toluene-d8	8.70	98	670311	50.12	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	100.24%		
62) 4-Bromofluorobenzene	11.12	95	233676	45.91	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.82%		

Target Compounds					Qvalue
16) Acetone	2.42	43	27049	16.909 ug/l	98
20) Methylene Chloride	2.83	84	33024	7.844 ug/l	99
43) Isopropyl Acetate	6.40	43	23082	2.871 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019846.D
Acq On : 11 Dec 2020 00:24
Operator : JC/MD
Sample : L4983-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR1-A-BOTTOM

Quant Time: Dec 11 03:12:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

