

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\  
 Data File : VX019847.D  
 Acq On : 11 Dec 2020 00:48  
 Operator : JC/MD  
 Sample : L4983-06  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 41 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 TR1-B-TOP

Quant Time: Dec 11 03:13:38 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X120420W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Dec 04 15:32:53 2020  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 5.63  | 168  | 313696   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.83  | 114  | 522923   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.09 | 117  | 474381   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.06 | 152  | 207662   | 50.00 | ug/l  | 0.00     |

| System Monitoring Compounds |                |     |          |       |        |      |
|-----------------------------|----------------|-----|----------|-------|--------|------|
| 33) 1,2-Dichloroethane-d4   | 6.03           | 65  | 200591   | 48.65 | ug/l   | 0.00 |
| Spiked Amount 50.000        | Range 61 - 141 |     | Recovery | =     | 97.30% |      |
| 35) Dibromofluoromethane    | 5.46           | 113 | 159321   | 47.14 | ug/l   | 0.00 |
| Spiked Amount 50.000        | Range 69 - 133 |     | Recovery | =     | 94.28% |      |
| 50) Toluene-d8              | 8.70           | 98  | 634663   | 49.19 | ug/l   | 0.00 |
| Spiked Amount 50.000        | Range 65 - 126 |     | Recovery | =     | 98.38% |      |
| 62) 4-Bromofluorobenzene    | 11.12          | 95  | 224709   | 45.76 | ug/l   | 0.00 |
| Spiked Amount 50.000        | Range 58 - 135 |     | Recovery | =     | 91.52% |      |

| Target Compounds       |      |    |       |             | Qvalue |
|------------------------|------|----|-------|-------------|--------|
| 16) Acetone            | 2.42 | 43 | 27222 | 17.566 ug/l | 99     |
| 20) Methylene Chloride | 2.83 | 84 | 33290 | 8.221 ug/l  | 99     |
| 43) Isopropyl Acetate  | 6.41 | 43 | 23981 | 3.092 ug/l  | 99     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\  
Data File : VX019847.D  
Acq On : 11 Dec 2020 00:48  
Operator : JC/MD  
Sample : L4983-06  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 41 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
TR1-B-TOP

Quant Time: Dec 11 03:13:38 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X120420W.M  
Quant Title : SW846 8260  
QLast Update : Fri Dec 04 15:32:53 2020  
Response via : Initial Calibration

