

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019848.D
 Acq On : 11 Dec 2020 01:11
 Operator : JC/MD
 Sample : L4983-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR1-B-BOTTOM

Quant Time: Dec 11 03:15:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	315110	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	527468	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	488340	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	216620	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.02	65	202140	48.81	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	97.62%	
35) Dibromofluoromethane	5.46	113	163238	47.89	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	95.78%	
50) Toluene-d8	8.70	98	643868	49.47	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	98.94%	
62) 4-Bromofluorobenzene	11.12	95	230441	46.52	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	93.04%	

Target Compounds					Qvalue
16) Acetone	2.42	43	27470	17.647 ug/l	97
20) Methylene Chloride	2.84	84	33089	8.120 ug/l	98
43) Isopropyl Acetate	6.41	43	24546	3.137 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019848.D
Acq On : 11 Dec 2020 01:11
Operator : JC/MD
Sample : L4983-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR1-B-BOTTOM

Quant Time: Dec 11 03:15:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

