

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019853.D
 Acq On : 11 Dec 2020 03:07
 Operator : JC/MD
 Sample : L4983-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR2-A-TOP

Quant Time: Dec 11 05:57:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	327104	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	554297	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	513299	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	218143	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	211762	49.26	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.52%		
35) Dibromofluoromethane	5.46	113	170547	47.61	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.22%		
50) Toluene-d8	8.70	98	682912	49.93	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	99.86%		
62) 4-Bromofluorobenzene	11.12	95	232094	44.59	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	89.18%		

Target Compounds

					Qvalue
16) Acetone	2.42	43	25767	15.946	ug/l 99
20) Methylene Chloride	2.84	84	32395	7.575	ug/l 99
43) Isopropyl Acetate	6.41	43	23198	2.821	ug/l 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019853.D
Acq On : 11 Dec 2020 03:07
Operator : JC/MD
Sample : L4983-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR2-A-TOP

Quant Time: Dec 11 05:57:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

