

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019856.D
 Acq On : 11 Dec 2020 04:17
 Operator : JC/MD
 Sample : L4983-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR2-B-BOTTOM

Quant Time: Dec 11 06:02:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	325339	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	541325	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	497820	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	211185	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	205300	48.01	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.02%
35) Dibromofluoromethane	5.46	113	164996	47.16	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.32%
50) Toluene-d8	8.70	98	662993	49.64	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.28%
62) 4-Bromofluorobenzene	11.12	95	230035	45.25	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.50%

Target Compounds					Qvalue
16) Acetone	2.42	43	25220	15.692	ug/l 98
20) Methylene Chloride	2.83	84	30974	7.225	ug/l 97
43) Isopropyl Acetate	6.41	43	22875	2.849	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019856.D
Acq On : 11 Dec 2020 04:17
Operator : JC/MD
Sample : L4983-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR2-B-BOTTOM

Quant Time: Dec 11 06:02:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

