

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019857.D
 Acq On : 11 Dec 2020 04:40
 Operator : JC/MD
 Sample : L4983-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR2-C-TOP

Quant Time: Dec 11 06:04:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	328779	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	539755	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	506670	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	216952	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	208449	48.24	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.48%		
35) Dibromofluoromethane	5.46	113	166458	47.72	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	95.44%		
50) Toluene-d8	8.70	98	667470	50.12	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	100.24%		
62) 4-Bromofluorobenzene	11.12	95	236017	46.56	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	93.12%		

Target Compounds					Qvalue
16) Acetone	2.42	43	25710	15.829 ug/l	97
20) Methylene Chloride	2.83	84	30303	6.948 ug/l	98
43) Isopropyl Acetate	6.41	43	21821	2.725 ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019857.D
Acq On : 11 Dec 2020 04:40
Operator : JC/MD
Sample : L4983-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR2-C-TOP

Quant Time: Dec 11 06:04:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration

