

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX121020\
 Data File : VX019856.D
 Acq On : 11 Dec 2020 04:17
 Operator : JC/MD
 Sample : L4983-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR2-B-BOTTOM

Quant Time: Dec 11 06:02:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 04 15:32:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	325339	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	541325	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	497820	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	211185	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	205300	48.01	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.02%		
35) Dibromofluoromethane	5.46	113	164996	47.16	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.32%		
50) Toluene-d8	8.70	98	662993	49.64	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	99.28%		
62) 4-Bromofluorobenzene	11.12	95	230035	45.25	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	90.50%		

Target Compounds


					Qvalue
16) Acetone	2.42	43	25220	15.692	ug/l 98
20) Methylene Chloride	2.83	84	30974	7.225	ug/l 97
43) Isopropyl Acetate	6.41	43	22875	2.849	ug/l 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX121020\
Data File : VX019856.D
Acq On : 11 Dec 2020 04:17
Operator : JC/MD
Sample : L4983-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR2-B-BOTTOM

Quant Time: Dec 11 06:02:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X120420W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 04 15:32:53 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 745

Delta R.T. 0.01 min

Lab File: VX019856.D

Acq: 11 Dec 2020 04:17

Instrument :

MSVOA_X

ClientSampleId :


TR2-B-BOTTOM

Tot Ion:168 Resp: 325339

Ion Ratio Lower Upper

168 100

99 57.2 45.8 68.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



5.63

100000

50000

0

Time-->



#16

Acetone

Concen: 15.692 ug/l

RT: 2.42 min Scan# 218

Delta R.T. 0.00 min

Lab File: VX019856.D


Acq: 11 Dec 2020 04:17

Tgt Ion: 43 Resp: 25220

Ion Ratio Lower Upper

43 100

58 36.1 27.8 41.8



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0

2.42


15000




10000

5000

0

Time-->





#20

Methylene Chloride

Concen: 7.225 ug/l

RT: 2.83 min Scan# 286

Delta R.T. 0.00 min

Lab File: VX019856.D

Acq: 11 Dec 2020 04:17

Instrument :

MSVOA_X

ClientSampleId :

TR2-B-BOTTOM

Tot Ion: 84 Resp: 30974





Ion Ratio Lower Upper

84 100

49 115.9 88.6 132.8

51 35.0 27.0 40.6

86 63.2 50.3 75.5



#33

1,2-Dichloroethane-d4

Concen: 48.013 ug/l

RT: 6.03 min Scan# 810

Delta R.T. 0.00 min

Lab File: VX019856.D


Acq: 11 Dec 2020 04:17


Tgt Ion: 65 Resp: 205300

Ion Ratio Lower Upper

65 100

67 53.5 0.0 105.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.83 min Scan# 942

Delta R.T. 0.01 min

Lab File: VX019856.D

Acq: 11 Dec 2020 04:17

Instrument :

MSVOA_X

ClientSampleId :

TR2-B-BOTTOM


Tot Ion:114 Resp: 541325

Ion Ratio Lower Upper

114 100

63 20.0 0.0 40.8




88 16.5 0.0 32.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 47.162 ug/l

RT: 5.46 min Scan# 718

Delta R.T. 0.01 min

Lab File: VX019856.D

Acq: 11 Dec 2020 04:17


Tgt Ion:113 Resp: 164996

Ion Ratio Lower Upper

113 100

111 104.1 81.9 122.9



192 20.5 16.7 25.1




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 2.849 ug/l

RT: 6.41 min Scan# 873

Delta R.T. 0.01 min

Lab File: VX019856.D

Acq: 11 Dec 2020 04:17

Instrument :

MSVOA_X

ClientSampleId :

TR2-B-BOTTOM





Tot Ion: 43 Resp: 22875

Ion Ratio Lower Upper

43 100

61 23.4 18.2 27.4

87 14.9 12.5 18.7



#50

Toluene-d8

Concen: 49.639 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. 0.00 min

Lab File: VX019856.D




Acq: 11 Dec 2020 04:17

Tgt Ion: 98 Resp: 662993

Ion Ratio Lower Upper

98 100

100 66.0 51.9 77.9





#62

4-Bromofluorobenzene

Concen: 45.250 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. 0.00 min

Lab File: VX019856.D

Acq: 11 Dec 2020 04:17

Instrument :

MSVOA_X

ClientSampleId :

TR2-B-BOTTOM


Tot Ion: 95 Resp: 230035

Ion Ratio Lower Upper

95 100

174 76.9 0.0 154.6

176 75.8 0.0 155.8






Abundance Ion 94.90 (94.60 to 95.60): VX019713.D (-1638) (-)

Ion 173.80 (173.50 to 174.50): VX019713.D (-1638) (-)

Ion 175.80 (175.50 to 176.50): VX019713.D (-1638) (-)

11.12

11.10 11.20



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.10 min Scan# 1478

Delta R.T. 0.01 min

Lab File: VX019856.D

Acq: 11 Dec 2020 04:17


Tgt Ion: 117 Resp: 497820

Ion Ratio Lower Upper

117 100

82 55.0 47.4 71.2

119 32.5 26.6 39.8





Abundance Ion 116.90 (116.60 to 117.60): VX019713.D (-1470) (-)


Ion 82.00 (81.70 to 82.70): VX019713.D (-1470) (-)

Ion 118.90 (118.60 to 119.60): VX019713.D (-1470) (-)

10.10

10.00 10.20





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.06 min Scan# 1800

Delta R.T. 0.00 min

Lab File: VX019856.D

Acq: 11 Dec 2020 04:17

Instrument :

MSVOA_X

ClientSampleId :

TR2-B-BOTTOM


Tot Ion:152 Resp: 211185

Ion Ratio Lower Upper

152 100

115 60.5 41.1 123.3

150 159.4 0.0 349.0



Abundance

Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

