

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121021\
 Data File : VX025708.D
 Acq On : 10 Dec 2021 17:58
 Operator : JC/MD
 Sample : VX1210WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK650

Quant Time: Dec 11 00:48:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 11 00:48:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	103308	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	95836	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	42885	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	32775	46.487	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.980%
7) Chloroethane-d5	1.666	69	29290	50.336	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.680%
11) 1,1-Dichloroethene-d2	2.306	63	47804	42.525	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.040%
21) 2-Butanone-d5	4.458	46	54011	102.913	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.910%
24) Chloroform-d	5.062	84	65205	52.918	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.840%
26) 1,2-Dichloroethane-d4	5.958	65	43875	57.126	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	114.260%
32) Benzene-d6	5.976	84	133265	52.467	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.940%
36) 1,2-Dichloropropane-d6	7.311	67	40248	51.180	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.360%
41) Toluene-d8	8.653	98	123055	51.328	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.660%
43) trans-1,3-Dichloroprop...	8.951	79	19276	48.754	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.500%
47) 2-Hexanone-d5	9.384	63	38018	92.873	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.870%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	56989	51.277	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.560%
66) 1,2-Dichlorobenzene-d4	12.323	152	45304	53.861	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121021\
Data File : VX025708.D
Acq On : 10 Dec 2021 17:58
Operator : JC/MD
Sample : VX1210WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK650

Quant Time: Dec 11 00:48:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 11 00:48:07 2021
Response via : Initial Calibration

