

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121021\
 Data File : VX025721.D
 Acq On : 10 Dec 2021 23:04
 Operator : JC/MD
 Sample : M4981-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0CP4

Quant Time: Dec 11 01:05:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM112221WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 11 00:48:07 2021
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John Carlone 12/11/2021
 Supervised By :Mahesh Dadoda 12/11/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	97136	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	90839	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	38729	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	27099	40.879	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	81.760%		
7) Chloroethane-d5	1.666	69	23701	43.319	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	86.640%		
11) 1,1-Dichloroethene-d2	2.306	63	40567	38.380	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	76.760%		
21) 2-Butanone-d5	4.458	46	45441	92.085	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	92.090%		
24) Chloroform-d	5.062	84	55213	47.656	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	95.320%		
26) 1,2-Dichloroethane-d4	5.964	65	36998	51.233	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	102.460%		
32) Benzene-d6	5.976	84	111975	46.511	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	93.020%		
36) 1,2-Dichloropropane-d6	7.312	67	33880	45.452	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	90.900%		
41) Toluene-d8	8.653	98	102465	45.091	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	90.180%		
43) trans-1,3-Dichloroprop...	8.951	79	15643	41.742	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	83.480%		
47) 2-Hexanone-d5	9.384	63	32184	82.947	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	82.950%		
56) 1,1,2,2-Tetrachloroeth...	11.195	84	48140	45.697	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	91.400%		
66) 1,2-Dichlorobenzene-d4	12.323	152	36322	47.817	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	95.640%		
Target Compounds					Qvalue	
12) 1,1-Dichloroethene	2.319	96	1172	2.247	ug/L #	1
20) cis-1,2-Dichloroethene	4.489	96	1055m	1.594	ug/L	
30) 1,1,1-Trichloroethane	5.385	97	819	0.817	ug/L #	60
34) Trichloroethene	7.129	95	4052	6.348	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121021\
Data File : VX025721.D
Acq On : 10 Dec 2021 23:04
Operator : JC/MD
Sample : M4981-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0CP4

Quant Time: Dec 11 01:05:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM112221WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 11 00:48:07 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 12/11/2021
Supervised By : Mahesh Dadoda 12/11/2021

