

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121024\
 Data File : VX044199.D
 Acq On : 10 Dec 2024 11:35
 Operator : JC/MD
 Sample : VX1210WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1210WBL01

Quant Time: Dec 11 01:27:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	110263	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	207593	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	178508	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	71476	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	94012	49.710	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.420%
35) Dibromofluoromethane	5.385	113	68829	46.401	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.800%
50) Toluene-d8	8.647	98	251203	49.128	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.260%
62) 4-Bromofluorobenzene	11.079	95	82733	47.543	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.080%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121024\
Data File : VX044199.D
Acq On : 10 Dec 2024 11:35
Operator : JC/MD
Sample : VX1210WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1210WBL01

Quant Time: Dec 11 01:27:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

