

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121024\
 Data File : VX044200.D
 Acq On : 10 Dec 2024 11:59
 Operator : JC/MD
 Sample : P5164-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Dec 11 01:28:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	112437	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	211530	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185283	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76755	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95552	49.548	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.100%		
35) Dibromofluoromethane	5.379	113	68775	45.501	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.000%		
50) Toluene-d8	8.647	98	255312	49.002	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.000%		
62) 4-Bromofluorobenzene	11.079	95	85535	48.238	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.480%		
Target Compounds						
3) Chloromethane	1.300	50	1780	1.188	ug/l #	78
16) Acetone	2.386	43	1228	1.386	ug/l	94
20) Methylene Chloride	2.782	84	872	0.580	ug/l #	47

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121024\
Data File : VX044200.D
Acq On : 10 Dec 2024 11:59
Operator : JC/MD
Sample : P5164-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Dec 11 01:28:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

