

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121024\
 Data File : VX044207.D
 Acq On : 10 Dec 2024 14:49
 Operator : JC/MD
 Sample : P5193-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Dec 11 01:34:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	111435	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	204955	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182775	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76353	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93575	48.959	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.920%
35) Dibromofluoromethane	5.385	113	67647	46.191	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.380%
50) Toluene-d8	8.647	98	251874	49.893	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.780%
62) 4-Bromofluorobenzene	11.079	95	86449	50.317	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.640%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121024\
Data File : VX044207.D
Acq On : 10 Dec 2024 14:49
Operator : JC/MD
Sample : P5193-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Dec 11 01:34:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

