

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121025\
 Data File : VX048801.D
 Acq On : 10 Dec 2025 11:47
 Operator : JC/MD
 Sample : Q3803-08DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CHASE-HDL

Quant Time: Dec 10 12:05:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	137775	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	283034	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	299189	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	162030	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	97166	52.545	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.100%
35) Dibromofluoromethane	5.373	113	88482	45.403	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.800%
50) Toluene-d8	8.635	98	316625	46.350	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.700%
62) 4-Bromofluorobenzene	11.061	95	137782	58.491	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.980%
Target Compounds						
16) Acetone	2.374	43	20501	28.960	ug/l	Qvalue 100
20) Methylene Chloride	2.788	84	27515	15.550	ug/l	99
40) Benzene	6.025	78	552178	69.785	ug/l	99
52) Toluene	8.702	92	223729	46.890	ug/l	100
67) Ethyl Benzene	10.177	91	22871	2.140	ug/l	99
68) m/p-Xylenes	10.281	106	42898	10.305	ug/l	98
69) o-Xylene	10.622	106	26303	6.743	ug/l	97
84) 1,2,4-Trimethylbenzene	11.732	105	20525	2.267	ug/l	87
95) Naphthalene	13.756	128	154592	17.107	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121025\
Data File : VX048801.D
Acq On : 10 Dec 2025 11:47
Operator : JC/MD
Sample : Q3803-08DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CHASE-HDL

Quant Time: Dec 10 12:05:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

