

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121025\
 Data File : VX048802.D
 Acq On : 10 Dec 2025 12:07
 Operator : JC/MD
 Sample : Q3803-09DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CHASE-IDL

Quant Time: Dec 10 12:27:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	145987	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	310496	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	345316	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	188614	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	123676	63.119	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 126.240%#		
35) Dibromofluoromethane	5.373	113	107123	50.107	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.220%		
50) Toluene-d8	8.635	98	369391	49.292	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 98.580%		
62) 4-Bromofluorobenzene	11.061	95	170213	65.868	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 131.740%#		
Target Compounds						
						Qvalue
16) Acetone	2.373	43	16483	21.974	ug/l	99
20) Methylene Chloride	2.794	84	13201	7.041	ug/l	92
40) Benzene	6.025	78	415595	47.878	ug/l	100
52) Toluene	8.702	92	147928	28.261	ug/l	98
67) Ethyl Benzene	10.177	91	13574	1.101	ug/l	97
68) m/p-Xylenes	10.281	106	24732	5.148	ug/l	100
69) o-Xylene	10.622	106	16091	3.574	ug/l	99
84) 1,2,4-Trimethylbenzene	11.732	105	11743	1.114	ug/l	79
95) Naphthalene	13.756	128	76458	6.312	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121025\
Data File : VX048802.D
Acq On : 10 Dec 2025 12:07
Operator : JC/MD
Sample : Q3803-09DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CHASE-IDL

Quant Time: Dec 10 12:27:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

