

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121025\
 Data File : VX048807.D
 Acq On : 10 Dec 2025 13:49
 Operator : JC/MD
 Sample : Q3787-07RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OWBR-02-170-120425RE

Quant Time: Dec 11 00:21:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	160525	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	337243	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	371573	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	199241	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	129478	60.095	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	120.200%#
35) Dibromofluoromethane	5.373	113	110759	47.699	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.400%
50) Toluene-d8	8.635	98	387316	47.585	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.160%
62) 4-Bromofluorobenzene	11.061	95	175644	62.579	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	125.160%#
Target Compounds						
16) Acetone	2.374	43	12860	15.592	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121025\
Data File : VX048807.D
Acq On : 10 Dec 2025 13:49
Operator : JC/MD
Sample : Q3787-07RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OWBR-02-170-120425RE

Quant Time: Dec 11 00:21:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

