

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121025\
 Data File : VX048808.D
 Acq On : 10 Dec 2025 14:10
 Operator : JC/MD
 Sample : Q3787-09RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OWBR-02-170-120425-FDRE

Quant Time: Dec 11 00:22:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	150728	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	313869	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	341495	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	184829	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	120292	59.461	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	118.920%#	
35) Dibromofluoromethane	5.379	113	102752	47.546	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	95.100%	
50) Toluene-d8	8.635	98	359487	47.455	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	94.900%	
62) 4-Bromofluorobenzene	11.061	95	159804	61.175	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	122.360%	
Target Compounds						
16) Acetone	2.374	43	13514	17.450	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121025\
Data File : VX048808.D
Acq On : 10 Dec 2025 14:10
Operator : JC/MD
Sample : Q3787-09RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OWBR-02-170-120425-FDRE

Quant Time: Dec 11 00:22:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

