

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121025\
 Data File : VX048810.D
 Acq On : 10 Dec 2025 14:51
 Operator : JC/MD
 Sample : Q3787-13RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OW-03B-51.5-120425-FDRE

Quant Time: Dec 11 00:23:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 05 03:27:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	152091	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	320261	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	354570	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	184719	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	125850	61.651	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 123.300%#	
35) Dibromofluoromethane	5.373	113	107682	48.833	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.660%	
50) Toluene-d8	8.635	98	375438	48.571	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.140%	
62) 4-Bromofluorobenzene	11.061	95	166905	62.619	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 125.240%#	
Target Compounds						
16) Acetone	2.380	43	3269	4.183	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX121025\
Data File : VX048810.D
Acq On : 10 Dec 2025 14:51
Operator : JC/MD
Sample : Q3787-13RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-03B-51.5-120425-FDRE

Quant Time: Dec 11 00:23:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120425W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 05 03:27:34 2025
Response via : Initial Calibration

